



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
09/990,932	11/21/2001	Robert J. Sicurelli JR.	CONTINUATION-11-D1	4222
4988	7590	08/01/2006	EXAMINER	
ALFRED M. WALKER 225 OLD COUNTRY ROAD MELVILLE, NY 11747-2712			WILSON, JOHN J	
			ART UNIT	PAPER NUMBER
			3732	

DATE MAILED: 08/01/2006

Please find below and/or attached an Office communication concerning this application or proceeding.

# Office Action Summary

Application No.

09/990,932

Applicant(s)

SICURELLI ET AL.

Examiner

John J. Wilson

Art Unit

3732

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

## Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

## Status

- 1) ☒ Responsive to communication(s) filed on 21 June 2006.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

## Disposition of Claims

- 4) ☒ Claim(s) 33-35,38-40,42,44-46,50,52-61,64,65 and 70-100 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 33-35,38-40,42,44-46,50,52-61,64,65 and 70-100 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

## Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

## Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some \* c) ☐ None of:
- 1) ☐ Certified copies of the priority documents have been received.
  - 2) ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  - 3) ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

## Attachment(s)

- 1) ☐ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) ☒ Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)  
Paper No(s)/Mail Date 6/21/06
- 4) ☐ Interview Summary (PTO-413)  
Paper No(s)/Mail Date. \_\_\_\_\_
- 5) ☐ Notice of Informal Patent Application (PTO-152)
- 6) ☐ Other: \_\_\_\_\_

**DETAILED ACTION**

***Claim Rejections - 35 USC § 103***

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 33, 35, 38, 42, 44-46, 50, 53-61, 64, 65, 70, 71, 74-82, 84, 85, 88, 89 and 91-100 are rejected under 35 U.S.C. 103(a) as being unpatentable over Reynaud et al (5328372) in view of Albert (5564929). Reynaud shows a post comprising a bundle of non-metallic and non-woven fibers 5 in a resin 4. The fibers and resin of Reynaud are inherently flexible to some degree, however, Reynaud shows using carbon fibers not glass. Albert teaches the use of alternative fibers including carbon or glass, column 2, lines 59-65. It would be obvious to one of ordinary skill in the art to modify Reynaud to include the use of glass fibers as suggested by Albert in order to make use of known alternative materials in order to obtain the desired known properties of those materials. Reynaud teaches matching properties of the tooth including the modulus of elasticity. To use the inherently more flexible glass to better match such properties would have been obvious to one of ordinary skill in the art. The specific shape of the post used is an obvious matter of choice in shape to best match the canal. The specific type of glass fibers used is an obvious matter of choice in known materials to one of ordinary skill in

Art Unit: 3732

the art. The specific type of resin used is an obvious matter of choice in the use of known materials to the skilled artisan. To include surface texturing or facets are well known to one of ordinary skill in the implant art in order to improve the hold in the bone and in order to position the implant in the bone. To call the post, a pin, is merely terminology, and therefore, is not given patentable weight. Reynaud shows compacted fibers in the drawings, however, does not state a type or degree of compacting. The limitation "loosely" is a relative term that describes a degree of compacting that can depend on comparison and/or interpretation, and as such, the degree of compactness of the fibers is an obvious matter of choice in the degree of a known parameter to one of ordinary skill in the art. Reynaud does not show twisted fibers. Albert shows twisted fibers in Figs. 4 and 8. It would be obvious to one of ordinary skill in the art to modify Reynaud to include twisting the fibers as shown by Albert in order to make use of known shapes to obtain the desired properties.

Claims 34 and 90 are rejected under 35 U.S.C. 103(a) as being unpatentable over Reynaud et al (5328372) in view of Albert (5564929) as applied to the claims above, and further in view of Kwiatkowski (4936776). The above combination does not show translucent. Kwiatkowski teaches using a translucent post. It would be obvious to one of ordinary skill in the art to modify the above combination to include a translucent post as shown by Kwiatkowski in order to preserve the normal look of a tooth.

Art Unit: 3732

Claim 39 is rejected under 35 U.S.C. 103(a) as being unpatentable over Reynaud et al (5328372) in view of Albert (5564929) as applied to the claims above, and further in view of Al Kasem (5326264). The above combination does not show using an opaque material. Al Kasem teaches using an opaque filler, column 18, line 19. It would be obvious to one of ordinary skill in the art to modify the above combination to include the use of an opaque material as shown by Al Kasem in order to make use of known materials for best matching the tooth.

Claims 40 and 52 are rejected under 35 U.S.C. 103(a) as being unpatentable over Reynaud et al (5328372) in view of Albert (5564929) as applied to the claims above, and further in view of Weissman (5326263). The above combination does not show a end shaped to be rounded and to direct light. Weissman shows an end shaped to direct light, Fig. 6, that can be rounded, column 5, lines 1-3 and column 6, lines 1-4, and to direct light. It would be obvious to one of ordinary skill in the art to modify the above combination to include a shaped end as shown by Weissman in order to direct light. To shape the end by polishing is an obvious matter of choice in the process used to obtain a known structure to the skilled artisan.

Claims 72 and 73 are rejected under 35 U.S.C. 103(a) as being unpatentable over Reynaud et al (5328372) in view of Albert (5564929) and Al Kasem (5326264) as applied to the claim 39 above, and further in view of Fujisawa et al (4931096). The above combination does not show the use of a radio opaque material. Fujisawa

teaches the use of radio opaque material including barium sulfate, column 2, lines 10-16. It would be obvious to one of ordinary skill in the art to modify the above combination to include radio opaque material as shown by Fujisawa in order to enable the material to show up on radiograph.

Claims 83, 86 and 87 are rejected under 35 U.S.C. 103(a) as being unpatentable over Reynaud et al (5328372) in view of Albert (5564929) as applied to the claims above, and further in view of Fujisawa et al (4931096). The above combination does not show the use of a radio opaque material. Fujisawa teaches the use of radio opaque material including barium sulfate, column 2, lines 10-16. It would be obvious to one of ordinary skill in the art to modify the above combination to include radio opaque material as shown by Fujisawa in order to enable the material to show up on radiograph.

#### ***Information Disclosure Statement***

The IDS filed June 21, 2006 has been considered and an initialed copy attached.

#### ***Response to Arguments***

Applicant's arguments filed June 21, 2006 have been fully considered but they are not persuasive. Applicant argues that the 5,518,399 patent teaches using fiberglass with the embodiment shown in Fig. 8, reasoning the language used, "spacer 20 and post reinforcing rod 30 are preferably formed from reinforced plastics such as fiberglass polyester composites", column 5, lines 53 and 54, in describing the embodiment of

Art Unit: 3732

Figs. 1-3, teaches that the language "reinforced plastics" includes fiberglass polyester composites, and as such, when the same terminology is used to describe the embodiment of Fig. 8, that is, that post 100 and rod 130 are made from "reinforced plastics", column 7, lines 35-37, this teaches using fiberglass polyester composites in this embodiment as well. This argument is disagreed with because the complete teaching of the material used in the embodiment of Figs. 1-3 is "reinforced plastics such as fiberglass polyester composites similar to those used in the construction of fishing poles". This teaches that the terminology "reinforced plastics" includes the fiberglass polyester composites used to make solid rods such as fishing poles and the rods described in the embodiment of Figs. 1-3. It would not be obvious, and in fact, it is not even clear how such a material would be used to make the bundle of fibers as shown in Fig. 8. As such, for the teaching of using fiberglass, the Albert reference is not pre-dated by the 5,518,399 patent. Applicant argues that Reynaud does not teach using another material. It is noted that Reynaud does teach using high strength fibers, see Abstract, while Alpert teaches that it is known to use glass fibers which is known to be a type of high strength fibers, and teaches using it in the same art. As such the combination is motivated and proper. With respect to twisted, Alpert shows this feature as described above.

***Conclusion***

Any inquiry concerning this communication or earlier communications from the examiner should be directed to John J. Wilson whose telephone number is 571-272-4722). The examiner can normally be reached on Monday through Thursday.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Cris Rodriguez, can be reached at 571-272-4964. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).



**John J. Wilson  
Primary Examiner  
Art Unit 3732**

jjw  
July 27, 2006